

HUMAN ADIPOCYTE FABP ELISA (HUMAN FABP4 ELISA)

Product Data Sheet

Cat. No.: RD191036200R

For Research Use Only

Page 1 of 28 ENG.001.A

CONTENTS

1.	INTENDED USE	3
2.	STORAGE, EXPIRATION	3
3.	INTRODUCTION	4
4.	TEST PRINCIPLE	5
5.	PRECAUTIONS	5
6.	TECHNICAL HINTS	6
7.	REAGENT SUPPLIED	6
8.	MATERIAL REQUIRED BUT NOT SUPPLIED	7
9.	PREPARATION OF REAGENTS	7
10.	PREPARATION OF SAMPLES	9
11.	ASSAY PROCEDURE	10
12.	CALCULATIONS	11
13.	PERFORMANCE CHARACTERISTICS	12
14.	DEFINITION OF THE STANDARD	17
15.	PRELIMINARY POPULATION AND CLINICAL DATA	17
16.	METHOD COMPARISON	20
17.	TROUBLESHOOTING AND FAQS	21
18.	REFERENCES	22
19.	EXPLANATION OF SYMBOLS	24

- This kit is manufactured by:
 Bio Vendor Research and Diagnostic Products, s.r.o.
- **>>** Use only the current version of Product Data Sheet enclosed with the kit!

Page 2 of 28 ENG.001.A

INTENDED USE

The RD191036200R Human AFABP ELISA is a sandwich enzyme immunoassay for the quantitative measurement of human AFABP.

Features

- It is intended for research use only
- The total assay time is less than 4 hours
- The kit measures total AFABP in serum and plasma (EDTA, citrate, heparin)
- Assay format is 96 wells
- Quality Controls are human serum based
- Standard is recombinant protein based
- Components of the kit are provided ready to use, concentrated or lyophilized
- Patent Application Number: DE 10 2005 034 788.6

2. STORAGE, EXPIRATION

Store the complete kit at 2-8°C. Under these conditions, the kit is stable until the expiration date (see label on the box).

For stability of opened reagents see Chapter 9.

Page 3 of 28 ENG.001.A

INTRODUCTION

Protein definition:

Protein name: Adipocyte FABP (AFABP)

Synonyms:

Fatty acid-binding protein, adipocyte (A-FABP)

Adipocyte lipid-binding protein (ALBP)

Gene name: FABP4 Swissprot: P15090 NCBI / Protein: P15090

Adipocyte fatty acid binding protein AFABP is a 15 kDa member of the intracellular fatty acid binding protein (FABP) family, which is known for the ability to bind fatty acids and related compounds (bile acids or retinoids) in an internal cavity. AFABP is expressed in a differentiation-dependent fashion in adipocytes and is a critical gene in the regulation of the biological function of these cells. In mice, targeted mutations in FABP4 (mouse gene is also called aP2 and its relevant protein P2 adipocyte protein or 3T3-L1 lipid binding protein) provide significant protection from hyperinsulinemia and insulin resistance in the context of both dietary and genetic obesity. Adipocytes obtained from AFABP-deficient mice also have reduced efficiency of lipolysis in vitro and in vivo, and these mice exhibited moderately improved systemic dyslipidemia. Recent studies also demonstrated AFABP expression in human macrophages upon differentiation and activation. In these cells, AFABP modulates inflammatory responses and cholesterol ester accumulation, and total or macrophage-specific AFABP deficiency confers dramatic protection against atherosclerosis in the apoE-/- mice. These results indicate a central role for AFABP in the development of major components of the metabolic syndrome through its distinct actions in adipocytes and macrophages.

Besides being active within the cell, AFABP appears to be a secreted protein (for normal levels and correlations with certain metabolic parameters see chapter 15). The extracellular role of secreted AFABP remains to be determined.

Areas of investigation:

Energy metabolism and body weight regulation

Page 4 of 28 ENG.001.A

4. TEST PRINCIPLE

In the BioVendor Human AFABP ELISA, standards, quality controls and samples are incubated in microplate wells pre-coated with polyclonal anti-human AFABP antibody. After 60 minutes incubation and washing, biotin labelled polyclonal anti-human AFABP antibody is added and incubated for 60 minutes with captured AFABP. After another washing, streptavidin-HRP conjugate is added. After 30 minutes incubation and the last washing step, the remaining conjugate is allowed to react with the substrate solution (TMB). The reaction is stopped by addition of acidic solution and absorbance of the resulting yellow product is measured. The absorbance is proportional to the concentration of AFABP. A standard curve is constructed by plotting absorbance values against concentrations of standards, and concentrations of unknown samples are determined using this standard curve.

PRECAUTIONS

- For professional use only
- Wear gloves and laboratory coats when handling immunodiagnostic materials
- Do not drink, eat or smoke in the areas where immunodiagnostic materials are being handled
- This kit contains components of human origin. These materials were found non-reactive for HBsAg, HCV antibody and for HIV 1/2 antigen and antibody. However, these materials should be handled as potentially infectious, as no test can guarantee the complete absence of infectious agents
- This kit contains components of animal origin. These materials should be handled as potentially infectious
- Avoid contact with the acidic Stop Solution and Substrate Solution, which contains hydrogen peroxide and tetramethylbenzidine (TMB). Wear gloves and eye and clothing protection when handling these reagents. Stop and/or Substrate Solutions may cause skin/eyes irritation. In case of contact with the Stop Solution and the Substrate Solution wash skin/eyes thoroughly with water and seek medical attention, when necessary
- The materials must not be pipetted by mouth

Page 5 of 28 ENG.001.A

6. TECHNICAL HINTS

- Reagents with different lot numbers should not be mixed
- Use thoroughly clean glassware
- Use deionized (distilled) water, stored in clean containers
- Avoid any contamination among samples and reagents. For this purpose, disposable tips should be used for each sample and reagent
- Substrate Solution should remain colourless until added to the plate. Keep Substrate Solution protected from light
- Stop Solution should remain colourless until added to the plate. The colour developed in the wells will turn from blue to yellow immediately after the addition of the Stop Solution. Wells that are green in colour indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution
- Dispose of consumable materials and unused contents in accordance with applicable national regulatory requirements

7. REAGENT SUPPLIED

Kit Components	State	Quantity
Antibody Coated Microtiter Strips	ready to use	96 wells
Biotin Labelled Antibody Conc. (100x)	concentrated	0.15 ml
Streptavidin-HRP Conjugate	ready to use	13 ml
Master Standard	lyophilized	1 vial
Quality Control HIGH	lyophilized	1 vial
Quality Control LOW	lyophilized	1 vial
Dilution Buffer	ready to use	2 x 13 ml
Biotin-Ab Diluent	ready to use	13 ml
Wash Solution Conc. (10x)	concentrated	100 ml
Substrate Solution	ready to use	13 ml
Stop Solution	ready to use	13 ml
Product Data Sheet + Certificate of Analysis	-	1 pc

Page 6 of 28 ENG.001.A

8. MATERIAL REQUIRED BUT NOT SUPPLIED

- Deionized (distilled) water
- Test tubes for diluting samples
- Glassware (graduated cylinder and bottle) for Wash Solution (Dilution Buffer)
- Precision pipettes to deliver 10-1000 μl with disposable tips
- Multichannel pipette to deliver 100 µl with disposable tips
- Absorbent material (e.g. paper towels) for blotting the microtitrate plate after washing
- Vortex mixer
- Orbital microplate shaker capable of approximately 300 rpm
- Microplate washer (optional). [Manual washing is possible but not preferable.]
- Microplate reader with 450 ± 10 nm filter, preferably with reference wavelength 630 nm (alternatively another one from the interval 550-650 nm)
- Software package facilitating data generation and analysis (optional)

9. PREPARATION OF REAGENTS

- All reagents need to be brought to room temperature prior to use
- Always prepare only the appropriate quantity of reagents for your test
- Do not use components after the expiration date marked on their label
- Assay reagents supplied ready to use:

Antibody Coated Microtiter Strips

Stability and storage:

Return the unused strips to the provided aluminium zip-sealed bag with desicant and seal carefully. Remaining Microtiter Strips are stable 3 months stored at 2-8°C and protected from the moisture.

Streptavidin-HRP Conjugate
Dilution Buffer
Biotin-Ab Diluent
Substrate Solution
Stop Solution

Stability and storage:

Opened reagents are stable 3 months when stored at 2-8°C.

Page 7 of 28 ENG.001.A

Assay reagents supplied concentrated or lyophilized:

Human AFABP Master Standard

Refer to the Certificate of Analysis for current volume of Dilution Buffer needed for reconstitution of standard!!!

Reconstitute the lyophilized Master Standard with Dilution Buffer just prior to the assay. Let it dissolve at least 15 minutes with occasional gentle shaking (not to foam). The resulting concentration of the human AFABP in the stock solution is **25 ng/ml**.

Prepare set of standards using Dilution Buffer as follows:

Volume of Standard	Dilution Buffer	Concentration
Stock	-	25 ng/ml
500 μl of stock	750 µl	10 ng/ml
500 μl of 10 ng/ml	500 μl	5 ng/ml
500 μl of 5 ng/ml	500 μl	2.5 ng/ml
500 μl of 2.5 ng/ml	750 µl	1 ng/ml
500 μl of 1 ng/ml	500 μl	0.5 ng/ml

Prepared Standards are ready to use, do not dilute them.

Stability and storage:

Standard stock solution (25 ng/ml) should be aliquoted and frozen at -20°C for 3 months. Avoid repeated freeze/thaw cycles.

Do not store the diluted Standard solutions.

Quality Controls HIGH, LOW

Refer to the Certificate of Analysis for current volume of distilled water needed for reconstitution and for current Quality Control concentration!!!

Reconstitute each Quality Control (HIGH and LOW) with distilled water just prior to the assay. Let it dissolve at least 15 minutes with occasional gentle shaking (not to foam).

Dilute reconstituted Quality Controls 10x with Dilution Buffer, e.g. 20 μ l of Quality Control + 180 μ l of Dilution Buffer when assaying samples in singlets, or preferably 30 μ l of Quality Control + 270 μ l of Dilution Buffer for duplicates.

Stability and storage:

The reconstituted Quality Controls must be used immediately or stored frozen at -20°C for 3 months. Avoid repeated freeze/thaw cycles.

Do not store the diluted Quality Controls.

Note:

Concentration of analyte in Quality Controls need not be anyhow associated with normal and/or pathological concentrations in serum or another body fluid. Quality Controls serve just for control that the kit works in accordance with PDS and CoA and that ELISA test was carried out properly.

Page 8 of 28 ENG.001.A

Biotin Labelled Antibody Conc. (100x)

Prepare the working Biotin Labelled Antibody solution by adding 1 part Biotin Labelled Antibody Concentrate (100x) with 99 parts Biotin-Ab Diluent. Example: 10 μ l of Biotin Labelled Antibody Concentrate (100x) + 990 μ l of Biotin-Ab Diluent for 1 strip (8 wells).

Stability and storage:

Opened Biotin Labelled Antibody Concentrate (100x) is stable 3 months when stored at 2-8°C. **Do not store the diluted Biotin Labelled Antibody solution.**

Wash Solution Conc. (10x)

Dilute Wash Solution Concentrate (10x) ten-fold in distilled water to prepare a 1x working solution. Example: 100 ml of Wash Solution Concentrate (10x) + 900 ml of distilled water for use of all 96-wells.

Stability and storage:

The diluted Wash Solution is stable 1 month when stored at 2-8°C. Opened Wash Solution Concentrate (10x) is stable 3 months when stored at 2-8°C.

10. PREPARATION OF SAMPLES

The kit measures AFABP in serum and plasma (EDTA, citrate, heparin) samples.

Samples should be assayed immediately after collection or should be stored at -20°C. Mix thoroughly thawed samples just prior to the assay and avoid repeated freeze/thaw cycles, which may cause erroneous results. Avoid using hemolyzed or lipemic samples.

Dilute serum or plasma samples 10x with Dilution Buffer just prior to the assay, e.g. 20 μ l of sample + 180 μ l of Dilution Buffer for singlets, or preferably 30 μ l of sample + 270 μ l of Dilution Buffer for duplicates. **Mix well** (not to foam). Vortex is recommended.

Stability and storage:

Samples should be stored at -20°, or preferably at -70°C for long-term storage. Avoid repeated freeze/ thaw cycles.

Do not store the diluted samples.

See Chapter 13 for stability of serum and plasma samples when stored at 2-8°C, effect of freezing/thawing and effect of sample matrix (serum/plasma) on the concentration of AFABP.

Note: It is recommended to use a precision pipette and a careful technique to perform the dilution in order to get precise results.

Page 9 of 28 ENG.001.A

11. ASSAY PROCEDURE

- 1. Pipet **100** μ**I** of diluted Standards, Quality Controls, Dilution Buffer (=Blank) and samples, preferably in duplicates, into the appropriate wells. See *Figure 1* for example of work sheet.
- 2. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.
- 3. Wash the wells 5-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
- 4. Add **100** μ I of Biotin Labelled Antibody into each well.
- 5. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.
- 6. Wash the wells 5-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
- 7. Add **100** μ**I** of Streptavidin-HRP Conjugate into each well.
- 8. Incubate the plate at room temperature (ca. 25°C) for **30 minutes**, shaking at ca. 300 rpm on an orbital microplate shaker.
- 9. Wash the wells 5-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
- 10. Add **100** μ**I** of Substrate Solution into each well. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with e.g. aluminium foil is recommended.
- 11. Incubate the plate for **10 minutes** at room temperature. The incubation time may be extended [up to 20 minutes] if the reaction temperature is below than 20°C. Do not shake the plate during the incubation.
- 12. Stop the colour development by adding **100** μ**I** of Stop Solution.
- 13. Determine the absorbance of each well using a microplate reader set to 450 nm, preferably with the reference wavelength set to 630 nm (acceptable range: 550-650 nm). Subtract readings at 630 nm (550-650 nm) from the readings at 450 nm. The absorbance should be read within 5 minutes following step 12.

Note: If the microplate reader is not capable of reading absorbance greater than the absorbance of the highest standard, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine AFABP concentration of off-scale samples. The readings at 405 nm should not replace the on-scale readings at 450 nm.

Note 2: Manual washing: Aspirate wells and pipet 0.35 ml Wash Solution into each well. Aspirate wells and repeat twice. After final wash, invert and tap the plate strongly against paper towel. Make certain that Wash Solution has been removed entirely.

Page 10 of 28 ENG.001.A

	strip 1+2	strip 3+4	strip 5+6	strip 7+8	strip 9+10	strip 11+12
Α	Standard 25	QC HIGH	Sample 8	Sample 16	Sample 24	Sample 32
В	Standard 10	Sample 1	Sample 9	Sample 17	Sample 25	Sample 33
С	Standard 5	Sample 2	Sample 10	Sample 18	Sample 26	Sample 34
D	Standard 2.5	Sample 3	Sample 11	Sample 19	Sample 27	Sample 35
Ε	Standard 1	Sample 4	Sample 12	Sample 20	Sample 28	Sample 36
F	Standard 0.5	Sample 5	Sample 13	Sample 21	Sample 29	Sample 37
G	Blank	Sample 6	Sample 14	Sample 22	Sample 30	Sample 38
Н	QC LOW	Sample 7	Sample 15	Sample 23	Sample 31	Sample 39

Figure 1: Example of a work sheet.

12. CALCULATIONS

Most microplate readers perform automatic calculations of analyte concentration. The Standard curve is constructed by plotting the mean absorbance (Y) of Standards against log of the known concentration (X) of Standards, using the four-parameter algorithm. Results are reported as concentration of AFABP ng/ml in samples.

Alternatively, the logit log function can be used to linearize the standard curve (i.e. logit of the mean absorbance (Y) is plotted against log of the known concentration (X) of Standards.

The measured concentration of samples and Quality Controls calculated from the standard curve must be multiplied by their respective dilution factor, because samples and Quality Controls have been diluted prior to the assay. e.g. 1.57 ng/ml (from standard curve) x 10 (dilution factor) = 15.7 ng/ml.

Page 11 of 28 ENG.001.A




Figure 2: Typical Standard Curve for Human AFABP ELISA.

13. PERFORMANCE CHARACTERISTICS

Typical analytical data of BioVendor Human AFABP ELISA are presented in this chapter.

Sensitivity

Limit of Detection (LOD) (defined as concentration of analyte giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: $A_{blank} + 3xSD_{blank}$) is calculated from the real AFABP values in wells and is 0.05 ng/ml.

*Dilution Buffer is pipetted into blank wells.

Limit of assay

Results exceeding AFABP level of 25 ng/ml should be repeated with more diluted samples. Dilution factor needs to be taken into consideration in calculating the AFABP concentration.

Page 12 of 28 ENG.001.A

Specificity

The antibodies used in this ELISA are specific for human AFABP.

Sera of several mammalian species were measured in the assay. See results below. For details please contact us at info@biovendor.com.

Mammalian serum	Observed
sample	crossreactivity
Bovine	no
Cat	no
Dog	yes
Goat	no
Hamster	no
Horse	no
Monkey	yes
Mouse	yes
Pig	no
Rabbit	no
Rat	no
Sheep	no

>> Presented results are multiplied by respective dilution factor

• **Precision**Intra-assay (Within-Run) (n=8)

Sample	Mean (ng/ml)	SD (ng/ml)	CV (%)
1	9.08	0.21	2.31
2	22.75	0.62	2.72

Inter-assay (Run-to-Run) (n=6)

Sample	Mean	SD	CV
	(ng/ml)	(ng/ml)	(%)
1	31.88	1.74	5.46
2	53.13	1.24	2.34

Page 13 of 28 ENG.001.A

• Spiking Recovery

Serum samples were spiked with different amounts of human AFABP and assayed.

Sample	O bserved	E xpected	Recovery O/E
	(ng/ml)	(ng/ml)	(%)
1	18.63	-	-
	69.43	68.63	101.2
	45.31	43.63	103.9
	31.87	28.63	111.3
2	23.44	-	-
	69.84	73.44	95.1
	52.55	48.44	108.5
	35.95	33.44	107.5

Linearity

Serum samples were serially diluted with Dilution Buffer and assayed.

Sample	Dilution	O bserved	E xpected	Recovery
		(ng/ml)	(ng/ml)	O/E (%)
1	-	59.50	-	-
	2x	29.90	29.75	100.5
	4x	14.41	14.87	96.8
	8x	7.57	7.44	101.7
2	-	52.62	-	-
	2x	26.37	26,31	100.2
	4x	13.01	13.15	98.9
	8x	6.77	6.58	102.9

Page 14 of 28 ENG.001.A

• Effect of sample matrix

Citrate, heparin and EDTA plasmas were compared to respective serum samples from the same 10 individuals.

Results are shown below:

Volunteer	Serum	PI	Plasma (ng/ml)		
No.	(ng/ml)	EDTA	Citrate	Heparin	
1	6.87	6.05	4.86	6.41	
2	13.51	12.18	10.93	12.16	
3	14.94	13.28	11.71	9.74	
4	12.67	12.04	11.30	12.55	
5	20.33	14.53	13.82	17.72	
6	10.46	9.43	8.94	9.68	
7	16.09	14.21	12.30	12.36	
8	13.94	12.30	11.90	13.94	
9	14.31	12.90	10.3	13.80	
10	7.55	6.92	6.64	7.42	
Mean (ng/ml)	13.07	11.38	10.33	11.58	
Mean Plasma/Serum	-	87.1	79.1	88.6	
(%)					
Coefficient of	-	0.91	0.84	0.91	
determination R ²					




Figure 3: AFABP levels measured using Human AFABP ELISA from 10 individuals using serum, EDTA, citrate and heparin plasma, respectively.

Page 15 of 28 ENG.001.A

Stability of samples stored at 2-8°C

Samples should be stored at -20° C. However, no decline in concentration of AFABP was observed in serum and plasma samples after 7 days when stored at 2-8°C. To avoid microbial contamination, samples were treated with ϵ -aminocaproic acid and thimerosal, resulting in the final concentration of 0.03% and 0.05%, respectively.

Sample	Incubation	Serum	PI	asma (ng	ı/ml)
	Temp, Period	(ng/ml)	EDTA	Citrate	Heparin
	-20°C	14.01	13.22	12.10	14.35
1	2-8°C, 1 day	13.93	12.75	10.88	13.48
	2-8°C, 7 days	14.05	12.98	11.12	13.44
	-20°C	5.40	3.96	3.84	5.67
2	2-8°C, 1 day	5.78	4.54	3.72	5.35
	2-8°C, 7 days	5.18	4.47	3.24	4.31
	-20°C	14.07	11.49	11.66	14.03
3	2-8°C, 1 day	14.21	11.68	11.49	12.63
	2-8°C, 7 days	13.56	11.59	11.43	12.73

• Effect of Freezing/Thawing

No decline was observed in concentration of human AFABP in serum and plasma samples after repeated (5x) freeze/thaw cycles. However it is recommended to avoid unnecessary repeated freezing/thawing of the samples.

Sample	Number of f/t	Serum	Plasma (ng/ml)		
Sample	cycles	(ng/ml)	EDTA	Citrate	Heparin
	1x	13.72	12.45	11.83	13.23
1	3x	12.96	12.71	10.22	13.27
	5x	11.71	12.06	11.57	12.37
	1x	18.87	16.00	15.35	17.77
2	3x	18.91	18.55	14.96	17.50
	5x	19.57	16.59	14.67	15.79
	1x	13.59	9.43	9.68	10.95
3	3x	11.71	10.37	8.22	11.09
	5x	11.67	10.62	10.78	10.86

Page 16 of 28 ENG.001.A

DEFINITION OF THE STANDARD

The recombinant protein is used as the standard in this assay. The recombinant AFABP is a 14.7 kDa protein containing 132 amino acid residues. Master Standard used in this kit contains 25 ng of AFABP measured by BCA method (Sigma-Aldrich).

15. PRELIMINARY POPULATION AND CLINICAL DATA

In our preliminary study, we investigated relations between serum AFABP value and some basic population parameters, parameters of lipid metabolism and parameters of insulin sensitivity.

Normal value and normal range in human serum

The mean value study with sera samples from 66 random selected women, 35-52 years old, has been established with the Biovendor Human AFABP ELISA in our laboratory (n=66, mean \pm SEM): **19.58 \pm 1.01 ng/ml.**

The normal range with sera samples from 66 random selected women, 35-52 years old, has been established with the Biovendor Human AFABP ELISA in our laboratory: Normal range (mean \pm 2 SD): **19.58** \pm **16.32 ng/ml**.

Reference range

The data quoted in these instructions should be used for guidance only. It is recommended that each laboratory include its own panel of control sample in the assay. Each laboratory should establish its own normal and pathological references ranges for AFABP levels with the assay.

Page 17 of 28 ENG.001.A




Figure 4: Histogram of serum AFABP values shows Gaussian distribution.

Concentration of AFABP (ng/ml)

• Influence of gender on AFABP value in human serum

The comparison of serum AFABP values between men (n=10) and women (n=26) selected by age (35-52 years old volunteers) did not show any significant difference. Results are presented as mean values:

	Age	Total	HDL	LDL	Triglycerides	AFABP
	(years)	cholesterol	(mmol/l)	(mmol/l)	(mmol/l)	(mmol/l)
	,	(mmol/l)	,	,	,	, ,
Women	45.5	5.21	1.46	3.35	1.13	21.28
Men	43.7	5.08	1.06	3.70	1.61	21.44

Page 18 of 28 ENG.001.A

Clinical relevance of serum AFABP value

We have found correlations between serum AFABP value and triglycerides, HDL and LDL values.




Figure 5: Serum AFABP positively correlates with triglycerides.




Figure 6: Serum AFABP negatively correlates with HDL.

Page 19 of 28 ENG.001.A




Figure 7: Serum AFABP positively correlates with LDL.

Increased serum AFABP values were found in obese volunteers in a similar study. AFABP values are presented as mean +/- SD:

	Volunteer No.	Age (years) AFABP		
	(F/M)		(ng/ml)	
Obese (BMI > 25)	129 (62/67)	53.9 ± 12.7	32.3 ± 14.8	
Non-obese (BMI < 25)	100 (46/54)	57.6 ± 12.8	20.0 ± 9.8	

16. METHOD COMPARISON

The BioVendor Human AFABP ELISA was not compared to the other commercial immunoassays.

Page 20 of 28 ENG.001.A

TROUBLESHOOTING AND FAQS

Weak signal in all wells

Possible explanations:

- Omission of a reagent or a step
- Improper preparation or storage of a reagent
- · Assay performed before reagents were allowed to come to room temperature
- Improper wavelength when reading absorbance

High signal and background in all wells

Possible explanations:

- Improper or inadequate washing
- Overdeveloping; incubation time with Substrate Solution should be decreased before addition of Stop Solution
- Incubation temperature over 30°C

High coefficient of variation (CV)

Possible explanation:

- Improper or inadequate washing
- Improper mixing Standards, Quality Controls or samples

Page 21 of 28 ENG.001.A

References to adipocyte FABP:

- Makowski L, Brittingham KC, Reynolds JM, Suttles J and Hotamisligil GS: The Fatty Acidbinding Protein, aP2, Coordinates Macrophage Cholesterol Trafficking and Inflammatory Activity. J Biol Chem. Apr 1; 280(13): 12888-12895 (2005)
- Maeda K, Cao H, Kono K, Gorgun CZ, Furuhashi M, Uysal KT, Cao Q, AtsumiG, Malone H, Krishnan B, Minokoshi Y, Kahn BB, Parker RA and Hotamisligil GS: Adipocyte/macrophage fatty acid binding proteins control integrated metabolic responses in obesity and diabetes. *Cell Metabolism*. February; 1(2): 107-119 (2005)
- Boord JB, Maeda K, Makowski L, Babaev VR, Fazio S, Linton MF, Hotamisligil GS: Combined adipocyte-macrophage fatty acid-binding protein deficiency improves metabolism, atherosclerosis, and survival in apolipoprotein E-deficient mice. *Circulation*. Sep 14; 110(11): 1492-1498 (2004)
- Lehmann F, Haile S, Axen E, Medina C, Uppenberg J, Svensson S, Lundback T, Rondahl L, Barf T: Discovery of inhibitors of human adipocyte fatty acid-binding protein, a potential type 2 diabetes target. *Bioorg Med Chem Lett*. Sep 6; **14**(17): 4445-4448 (2004)
- Damcott CM, Moffett SP, Feingold E, Barmada MM, Marshall JA, Hamman RF, Ferrell RE: Genetic variation in fatty acid-binding protein-4 and peroxisome proliferator-activated receptor gamma interactively influence insulin sensitivity and body composition in males. *Metabolism*. Mar; 53(3): 303-9 (2004)
- Jenkins-Kruchten AE, Bennaars-Eiden A, Ross JR, Shen WJ, Kraemer FB, Bernlohr DA: Fatty acid-binding protein-hormone-sensitive lipase interaction. Fatty acid dependence on binding. *J Biol Chem.* Nov 28; **278**(48): 47636-47643 (2003)
- Hertzel AV, Bennaars-Eiden A, Bernlohr DA: Increased lipolysis in transgenic animals overexpressing the epithelial fatty acid binding protein in adipose cells. *J Lipid Res.* Dec; 43(12): 2105-2111 (2002)
- Fu Y, Luo N, Lopes-Virella MF, Garvey WT: The adipocyte lipid binding protein (ALBP/aP2) gene facilitates foam cell formation in human THP-1 macrophages. *Atherosclerosis*. Dec; **165**(2): 259-269 (2002)
- Fisher RM, Hoffstedt J, Hotamisligil GS, Thorne A, Ryden M: Effects of obesity and weight loss on the expression of proteins involved in fatty acid metabolism in human adipose tissue. *Int J Obes Relat Metab Disord.* Oct; **26**(10): 1379-1385 (2002)
- Boord JB, Maeda K, Makowski L, Babaev VR, Fazio S, Linton MF, Hotamisligil GS: Adipocyte fatty acid-binding protein, aP2, alters late atherosclerotic lesion formation in severe hypercholesterolemia. *Arterioscler Thromb Vasc Biol.* Oct 1; 22(10): 1686-1691(2002)
- Scheja L, Makowski L, Uysal KT, Wiesbrock SM, Shimshek DR, Meyers DS, Morgan M, Parker RA, Hotamisligil GS: Altered insulin secretion associated with reduced lipolytic efficiency in aP2-/- mice. *Diabetes*. Oct; 48(10): 1987-1994 (1999)

Page 22 of 28 ENG.001.A

References to this product:

- Milner KL, van der Poorten D, Xu A, Bugianesi E, Kench JG, Lam KS, Chisholm DJ, George J. Adipocyte fatty acid binding protein levels relate to inflammation and fibrosis in nonalcoholic fatty liver disease. *Hepatology*. Jun; 49(6):1926-34 (2009)
- Yeung DC, Xu A, Tso AW, Chow WS, Wat NM, Fong CH, Tam S, Sham PC, Lam KS. Circulating levels of adipocyte and epidermal fatty acid-binding proteins in relation to nephropathy staging and macrovascular complications in type 2 diabetic patients. *Diabetes Care*. Jan; 32(1):132-4 (2009)
- Stejskal D, Karpisek M, Bronsky J. Serum adipocyte-fatty acid binding protein discriminates patients with permanent and temporary body weight loss. *J Clin Lab Anal.*; 22(5):380-2 (2008)
- Karpisek M, Stejskal D, Kotolova H, Kollar P, Janoutova G, Ochmanova R, Cizek L, Horakova D, Yahia RB, Lichnovska R, Janout V. Treatment with atorvastatin reduces serum adipocyte-fatty acid binding protein value in patients with hyperlipidaemia. *Eur J Clin Invest*. Aug; 37(8):637-42 (2007)
- Tso AW, Xu A, Sham PC, Wat NM, Wang Y, Fong CH, Cheung BM, Janus ED, Lam KS. Serum adipocyte fatty acid binding protein as a new biomarker predicting the development of type 2 diabetes: a 10-year prospective study in a Chinese cohort. *Diabetes Care*. Oct; 30(10):2667-72 (2007)
- Yeung DC, Xu A, Cheung CW, Wat NM, Yau MH, Fong CH, Chau MT, Lam KS. Serum adipocyte fatty acid-binding protein levels were independently associated with carotid atherosclerosis. *Arterioscler Thromb Vasc Biol.* Aug; 27(8):1796-802 (2007)
- Xu A, Tso AW, Cheung BM, Wang Y, Wat NM, Fong CH, Yeung DC, Janus ED, Sham PC, Lam KS: Circulating adipocyte-fatty acid binding protein levels predict the development of the metabolic syndrome: a 5-year prospective study. *Circulation*. Mar 27; **115**(12): 1537-1543 (2007)
- Bronsky J, Karpisek M, Bronska E, Pechova M, Jancikova B, Kotolova H, Stejskal D, Prusa R, Nevoral J: Adiponectin, adipocyte fatty acid binding protein, and epidermal fatty acid binding protein: Proteins newly identified in human breast milk. *Clin Chem.* Sep; 52(9): 1763-1770 (2006)
- von Eynatten M, Schneider JG, Allolio B: Adipocyte fatty acid binding protein (A-FABP) ist ein neuer Serum-Marker für Adipositas, Insulinresistenz und Metabolisches Syndrom. Diabetologie. 1: S23-S172 (2006)
- Xu A, Wang Y, Xu JY, Stejskal D, Tam S, Zhang J, Wat NM, Wong WK, Lam KS: Adipocyte fatty acid–binding protein is a plasma biomarker closely associated with obesity and metabolic syndrome. *Clin Chem.* Mar; **52**(3): 405-413 (2006)
- Stejskal D and Karpisek M: Adipocyte fatty acid binding protein (A-FABP) in Caucasian population: a new marker of metabolic syndrome? Eur J Clin Invest. Sep; 36(9): 621-625 (2006)

For more references on this product see our WebPages at www.biovendor.com


Page 23 of 28 ENG.001.A

19. EXPLANATION OF SYMBOLS

REF	Catalogue number
Cont.	Content
LOT	Lot number
<u>^!</u>	See instructions for use
	Biological hazard
	Expiry date
2 °C 1 8 °C	Storage conditions
25 PP	Identification of packaging materials

Page 24 of 28 ENG.001.A

Assay Procedure Summary




Page 25 of 28 ENG.001.A

	V	В	ပ	D	ш	ш	g	Ŧ
-								
2								
8								
4								
2								
9								
7								
8								
6								
10								
11								
12								

Page 26 of 28 ENG.001.A

Page 27 of 28 ENG.001.A





HEADQUARTERS: BioVendor - Laboratorní medicína a.s.	Karasek 1767/1	621 00 Brno CZECH REPUBLIC	Phone: Fax:	+420-549-124-185 +420-549-211-460	E-mail: Web:	info@biovendor.com sales@biovendor.com www.biovendor.com
EUROPEAN UNION: BioVendor GmbH	Im Neuenheimer Feld 583	D-69120 Heidelberg GERMANY		+49-6221-433-9100 +49-6221-433-9111	E-mail:	infoEU@biovendor.com
USA, CANADA AND MEXICO: BioVendor LLC	128 Bingham Rd. Suite 1300	Asheville, NC 28806 USA	Phone: Fax:	+1-828-575-9250 +1-800-404-7807 +1-828-575-9251	E-mail:	infoUSA@biovendor.com
CHINA - Hong Kong Office: BioVendor Laboratories Ltd	Room 4008 Hong Kong Plaza, No.188	Connaught Road West Hong Kong, CHINA		+852-2803-0523 +852-2803-0525	E-mail:	infoHK@biovendor.com
CHINA – Mainland Office: BioVendor Laboratories Ltd	Room 2917, 29/F R & F Ying Feng Plaza, No.2 Huaqiang road	Pearl River New Town Guang Zhou, CHINA		+86-20-38065519 +86-20-38065529	E-mail:	infoCN@biovendor.com

Page 28 of 28 ENG.001.A